

**Tannin**

Code : XL-209

Range : 5 – 100 &amp; 50 – 1,000 ppm as Tannin.

**AQUA-XL**  
Water Analysing Kits**Directions for use :**

1. Take 10 ml of water sample to be tested in the Test jar.
  2. Add 2-3 micro spoon full of Reagent TN-1.
  3. Mix contents well to dissolve.
  4. Now add Reagent TN-2 drop wise, counting the number of drops while mixing until **the PINK colour appears & which persists for 1 minute.**
- # If expected ppm level of Tannin is more than 100 ppm then use Reagent TN-3 instead of TN-2.

**Calculations :**

Tannin as ppm

$$= 5 \times \text{Number of drops of Reagent TN-2.}$$

$$= 50 \times \text{Number of drops of Reagent TN-3.}$$